

Determining “Born-On” Date for Fyrite Fluids



- In the example above, the Fyrite Fluid was “born” on Nov. 1st, 2017.
- The 2 55 1 is a batch number that has only internal usage for Bacharach traceability.
- CO2 Fyrite Fluid has ~ 3 years, ~300 tests before expiration.
- O2 Fyrite Fluid has ~1 year, ~100 tests before expiration.
- The accumulation of particulates in the fluid signals the fluid may be near expiration.